

Quick Start Guide How to Use Your Chamber

Remove inhaler and mouthpiece caps.



Shake the inhaler.



Insert the inhaler into the back of the chamber.



Apply mouthpiece or mask to face.






Monaghan Medical Corporation Plattsburgh, NY 12901

518.561.7330 phone 800.833.9653 toll free 518.561.5088 fax

monaghanmed.com

AEROCHAMBER PLUS® Z STAT® aVHC

People who use a valved holding chamber with their inhaler have better control.¹



Press the inhaler, breathe in slowly and hold your breath for 5 to 10 seconds or breathe 2 to 3 times.

(For mask, count for 5 to 6 breaths.)

Repeat steps 2 to 5 as prescribed.





AEROCHAMBER PLUS® Z STAT® aVHC



Quick Start Guide Cleaning Your Chamber

Remove the back piece only.


2 Soak 15 minutes in soapy water, agitate gently and rinse or place in dishwasher on top rack.



Dishwasher instructions can be found with the complete instructions packaged with your AEROCHAMBER PLUS® Z STAT® aVHC.

3 Shake out water and dry vertically.

Reassemble the chamber.



These are abbreviated use and cleaning instructions, please remember to read the complete instructions that are packaged with your AEROCHAMBER PLUS® Z STAT® aVHC device.



Monaghan Medical Corporation Plattsburgh, NY 12901

518.561.7330 phone 800.833.9653 toll free 518.561.5088 fax



Words or phrases accompanied by ™ and ® are trademarks and registered trademarks of Monaghan Medical Corporation or an affiliate of Monaghan Medical Corporation, © 2015-2019 Monaghan Medical Corporation, † trademarks and registered trademarks of the respective companies. 1. Levy et al. Prim Care Respir J. 2013;22(4):406-11. P/N 79250-06, 09/2019